

Australian Standard™

**Medical equipment—Conical fittings
with a 6 percent (Luer) taper for
syringes, needles and certain other
medical equipment**

Part 2: Lock fittings

This Australian Standard was prepared by Committee HE-009, Hypodermic Equipment - General Medical. It was approved on behalf of the Council of Standards Australia on 16 May 2003 and published on 30 June 2003.

The following are represented on Committee HE-009:

Australian Chamber of Commerce and Industry
Australian College of Operating Room Nurses
Australian Infection Control Association
Australian Medical Association
Certification Bodies (Australia)
Commonwealth Department of Health and Ageing
Department of Human Services (Victoria)
Federation of Sterilizing Research and Advisory Councils of Australia
Hunter Area Health Service
Medical Industry Association of Australia Inc
NSW Health Department
Queensland Health
Royal Australian College of General Practitioners
Royal College of Pathologists of Australasia
Royal Melbourne Hospital

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 03167.

Australian Standard™

Medical equipment—Conical fittings with a 6 percent (Luer) taper for syringes, needles and certain other medical equipment

Part 2: Lock fittings

First published as AS 1600.2—2003.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 5327 2

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-009, Hypodermic Equipment—General Medical. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with and has been reproduced from ISO 594-2:1998, *Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment—Part 2: Lock fittings*.

The objective of this Standard is to specify requirements for conical lock fittings with a 6 percent (Luer) taper for use with hypodermic syringes and needles and with certain other apparatus for medical use, e.g. transfusion equipment, for fittings made of rigid and of semi-rigid materials and include test methods.

The terms ‘normative’ and ‘informative’ are used to define the application of the annex to which they apply. A normative annex is an integral part of a standard, whereas an informative annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS	
594	Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment	1600	Medical equipment—Conical fittings with a 6 percent (Luer) taper for syringes, needles and certain other medical equipment
594-1	Part 1: General requirements	1600.1	Part 1: General requirement

CONTENTS

1 Scope	1
2 Normative references	1
3 Dimensions and tolerances	1
4 Requirements.....	6
5 Test methods	6

NOTES

AUSTRALIAN STANDARD

Medical equipment—Conical fittings with a 6 percent (Luer) taper for syringes, needles and certain other medical equipment

Part 2: Lock fittings

1 Scope

This part of ISO 594 specifies requirements for conical lock fittings with a 6 % (Luer) taper for use with hypodermic syringes and needles and with certain other apparatus for medical use, e.g. transfusion equipment.

The requirements apply to fittings made of rigid and of semi-rigid materials and include test methods, but exclude provision for more flexible or elastomeric materials.

NOTE 1 It is not practicable to define the characteristics of rigid or semi-rigid materials with precision, but glass and metal may be considered as typical rigid materials. In contrast many plastics materials may be regarded as semi-rigid.

NOTE 2 The Luer lock fitting was designed for use at pressures of the order of 300 kPa or lower. Its use in other applications may require consideration to establish its suitability.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 594. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 594 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 468, *Surface roughness — Parameters, their values and general rules for specifying requirements.*

ISO 594-1:1986, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements.*

ISO 7886-1, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use.*

3 Dimensions and tolerances

3.1 Male and female 6 % (Luer) conical fittings

The dimensions and tolerances for the male and female fittings specified in ISO 594-1 apply to the relevant conical part of the fitting described in clause 4 of this part of ISO 594.

3.2 Male and female 6 % (Luer) conical lock fittings

3.2.1 Rigid materials

The dimensions of male and female lock fittings made of rigid materials shall be as shown in Figures 1 to 4 and as given in Table 1.

This is a free preview. Purchase the entire publication at the link below:

-
- ▶ Looking for additional Standards? Visit [SAI Global Infostore](#)
 - ▶ Subscribe to our [Free Newsletters about Australian Standards® in Legislation; ISO, IEC, BSI and more](#)
 - ▶ Do you need to [Manage Standards Collections Online?](#)
 - ▶ Learn about [LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
 - ▶ Do you want to [know when a Standard has changed?](#)
 - ▶ Want to [become an SAI Global Standards Sales Affiliate?](#)

Learn about other SAI Global Services:

- ▶ [LOGICOM Military Parts and Supplier Database](#)
- ▶ [Metals Infobase Database of Metal Grades, Standards and Manufacturers](#)
- ▶ [Materials Infobase Database of Materials, Standards and Suppliers](#)
- ▶ [Database of European Law, CELEX and Court Decisions](#)

Need to speak with a Customer Service Representative - [Contact Us](#)